



Advanced Analytical Solutions *Your laboratory evaluation solution.*



www.AdvancedQA.com
Info@AdvancedQA.com

6325 Emerson Avenue
Parkersburg, WV 26104

(800) 283-4844
(304) 485-6325

SAMPLE PREPARATION INSTRUCTIONS

BENCHMARK PERFORMANCE AUDIT WASTE WATER (WP) SERIES

Please follow these guidelines for preparing your sample. Proper sample preparation along with appropriate QC methods will allow the greatest chance for assay success. All samples should be mixed well before use and stored at 4°C upon dilution.

Volatile Solids:

Using a Class A Pipet, transfer 10.0 ml from the vial to a Class A one (1) liter volumetric flask, and dilute to the mark with laboratory grade water. **DO NOT FILTER THE SAMPLE.** Just analyze the normal volume that you would run.